







**health**

Department:  
Health  
PROVINCE OF KWAZULU-NATAL

E3099886

ZNQ 628/20/21

ZNQSI TAL 179/21/22

**Specifications Template**

Prepared by:

Initial and Surname	Designation	Signature	Date
M Nyuswa	Buyer		12-03-21

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
A.N. Mncube	Om		12-03-2021

Item details	Specification
Item description	Catheter, Polyetherane, Radio-Opaque For Central Arterial and Peripheral Catheterization.
Size	2FR X4cm polyetherane catheter 22G X42mm Introducer needle 0.46mm X230mm straight guidewire with flexible tip 40-50cm X45-50cm fenestrated drape 4 X ball swabs
Colour	
Material	
Packaging (unit/box)	Each( set, sterile peel back pack)
Functionality/performance	Must be easy to insert and the care and maintenance of line must be simple
Purpose	
Other:	

Approved by specifications committee chairperson:

Initial and Surname	Portfolio	Signature	Date
N.P Mthembu	SCM		13.03.2021