



health
 Department:
 Health
 PROVINCE OF KWAZULU-NATAL

ESISS368

Specifications Template

Prepared by:

Initial and Surname	Designation	Signature	Date
C Mhlongo	Buyer		16.02.2021

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
JZ Shabale	Pa		17-02-21

Item details	Specification
Item description	FILTER, SYRINGE
Size	0.22 µm pore size/ 33MM diameter
Colour	WHITE AND BLUE
Material	Individual packed, Sterile Mixed Cellulose Esters membrane
Packaging (unit/box)	EACH
Functionality/performance	Purity
Purpose	1 To filter gas during Eye surgery. 2 To prevent air entry during insertion ICP MONITOR in NEUROSURGERY
Other:	

Approved by specifications committee chairperson:

Initial and Surname	Portfolio	Signature	Date
N. P. Mthembu	Scit		18.02.21