

E3006225



health

Department:
Health
PROVINCE OF KWAZULU-NATAL

E3006225

Prepared by:

Initial and Surname	Designation	Signature	Date

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
B Kusel	Anaesthesia specialist		18/08/2021

Item description	Intravenous cannula with introducer needle with injection port and fixation wings
Size	Size: 16 G Length: 45 mm \pm 2 mm OD: 1.7 mm Flow rate: 170 \pm 5%
Colour	Grey
Material	Cannula manufactured from teflon / polytetrafluoroethylene/ polyurethane - radio-opaque Needle manufactured from stainless steel
Packaging (unit/box)	Box
Functionality/performance	Triple bevelled needle through cannula. Needle must be flush with a tapered cannula to pass easily through the skin. Cannula must not flange and must be kink resistant. Cannula to advance easily over needle – must have advancing hub/pusher Cannula to re-insert and re-engage over needle easily whilst in use. Hub must be translucent so that flash-back can be easily observed
Purpose	Intravenous access for fluid and drug administration
Other:	Individually packed in peel pouch with view paper For single use only

Approved by specifications committee chairperson:

Initial and Surname	Portfolio	Signature	Date
N.P Mthembu	Scup		18 08. 2021

E3006227



health

Department:
Health
PROVINCE OF KWAZULU-NATAL

E3006227

Prepared by:

Initial and Surname	Designation	Signature	Date

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
B Kusol	Anaesthesia specialist		18/08/2021

Item description	
Item description	Intravenous cannula with introducer needle with injection port and fixation wings
Size	Size: 18 G Length: 30 mm ± 2 mm OD: 1.3 mm Flow rate: 105ml/min ± 5%
Colour	Green
Material	Cannula manufactured from teflon / polytetrafluoroethylene/ polyurethane - radio-opaque Needle manufactured from stainless steel
Packaging (unit/box)	Box
Functionality/performance	Triple bevelled needle through cannula. Needle must be flush with a tapered cannula to pass easily through the skin. Cannula must not flange and must be kink resistant. Cannula to advance easily over needle – must have advancing hub/pusher Cannula to re-insert and re-engage over needle easily whilst in use. Hub must be translucent so that flashback can be easily observed
Purpose	Intravenous access for fluid and drug administration
Other:	Individually packed in peel pouch with view paper For single use only
N.P. Mthembu SAMP 18.08.21	

E3006228



health

Department:
Health
PROVINCE OF KWAZULU-NATAL

E3006228

Prepared by:

Initial and Surname	Designation	Signature	Date

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
B Kusel	Anaesthesia specialist		18/08/2021

Item description	
Item description	Intravenous cannula with introducer needle with injection port and fixation wings
Size	Size: 20 G Length: 30 mm \pm 2 mm OD: 1.1 mm Flow rate: 60 ml/min \pm 5%
Colour	Pink
Material	Cannula manufactured from teflon / polytetrafluoroethylene/ polyurethane - radio-opaque Needle manufactured from stainless steel
Packaging (unit/box)	Box
Functionality/performance	Triple bevelled needle through cannula. Needle must be flush with a tapered cannula to pass easily through the skin. Cannula must not flange and must be kink resistant. Cannula to advance easily over needle - must have advancing hub/pusher Cannula to re-insert and re-engage over needle easily whilst in use. Hub must be translucent so that flashback can be easily observed
Purpose	Intravenous access for fluid and drug administration
Other:	Individually packed in peel pouch with view paper For single use only
Nolwazi Mthembu SCMPact	
18.08.2021	

E3006229



health

Department:
Health
PROVINCE OF KWAZULU-NATAL

E3006229

Prepared by:

Initial and Surname	Designation	Signature	Date

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
B Kusel	Anaesthesia specialist		18/08/2021

Item description	
Item description	Intravenous cannula with introducer needle with injection port and fixation wings
Size	Size: 22 G Length: 25 mm \pm 2 mm OD: 0.90 mm Flow rate: 38 ml/min \pm 5%
Colour	Blue
Material	Cannula manufactured from teflon / polytetrafluoroethylene/ polyurethane - radio-opaque Needle manufactured from stainless steel
Packaging (uni/box)	Box
Functionality/performance	Triple bevelled needle through cannula. Needle must be flush with a tapered cannula to pass easily through the skin. Cannula must not flange and must be kink resistant. Cannula to advance easily over needle - must have advancing hub/pusher Cannula to re-insert and re-engage over needle easily whilst in use. Hub must be translucent so that flash-back can be easily observed
Purpose	Intravenous access for fluid and drug administration
Other:	Individually packed in peel pouch with view paper For single use only
N.P. Mthembu Sompact	18.08.2021

DECLARATION OF INTEREST

1. Any legal person, including persons employed by the state¹, or persons having a kinship with persons employed by the state, including a blood relationship, may make an offer or offers in terms of this invitation to quote (includes a price quotation, advertised competitive quote, limited quote or proposal). In view of possible allegations of favouritism, should the resulting quote, or part thereof, be awarded to persons employed by the state, or to persons connected with or related to them, it is required that the bidder or his/her authorised representative declare his/her position in relation to the evaluating/adjudicating authority where-
 - the bidder is employed by the state; and/or
 - the legal person on whose behalf the bidding document is signed, has a relationship with persons/a person who are/is involved in the evaluation and or adjudication of the quote(s), or where it is known that such a relationship exists between the person or persons for or on whose behalf the declarant acts and persons who are involved with the evaluation and or adjudication of the quote.
2. In order to give effect to the above, the following questionnaire must be completed and submitted with the quote.

- | | |
|--|---|
| 2.1. Full Name of bidder/representative..... | 2.4. Company Registration Number: |
| 2.2. Identity Number: | 2.5. Tax Reference Number: |
| 2.3. Position occupied in the Company (director, trustee, shareholder ²):..... | 2.6. VAT Registration Number: |

2.7. The names of all directors / trustees / shareholders / members, their individual identity numbers, tax reference numbers and, if applicable, employee / persal numbers must be indicated in paragraph 3 below. TICK APPLICABLE

2.8. Are you or any person connected with the bidder presently employed by the state? YES NO

2.8.1. If so, furnish the following particulars:
 Name of person / director / trustee / shareholder/ member:
 Name of state institution at which you or the person connected to the bidder is employed:.....
 Position occupied in the state institution: Any other particulars:.....

2.8.2. If you are presently employed by the state, did you obtain the appropriate authority to undertake remunerative work outside employment in the public sector? YES NO

2.8.2.1. If yes, did you attach proof of such authority to the quote document?

(Note: Failure to submit proof of such authority, where applicable, may result in the disqualification of the quote.)

2.8.2.2. If no, furnish reasons for non-submission of such proof:

2.9. Did you or your spouse, or any of the company's directors / trustees / shareholders / members or their spouses conduct business with the state in the previous twelve months? YES NO

2.9.1. If so, furnish particulars:.....

2.10. Do you, or any person connected with the bidder, have any relationship (family, friend, other) with a person employed by the state and who may be involved with the evaluation and or adjudication of this quote? YES NO

2.10.1. If so, furnish particulars:.....

2.11. Are you, or any person connected with the bidder, aware of any relationship (family, friend, other) between any other bidder and any person employed by the state who may be involved with the evaluation and or adjudication of this quote? YES NO

2.11.1. If so, furnish particulars:.....

2.12. Do you or any of the directors / trustees / shareholders / members of the company have any interest in any other related companies whether or not they are bidding for this contract? YES NO

2.12.1. If so, furnish particulars:.....

3. Full details of directors / trustees / members / shareholders.

NB: The Department Of Health will validate details of directors / trustees / members / shareholders on CSD. It is the suppliers' responsibility to ensure that their details are up-to-date and verified on CSD. If the Department cannot validate the information on CSD, the quote will not be considered and passed over as non-compliant according to National Treasury Instruction Note 4 (a) 2016/17.

4 DECLARATION

I, THE UNDERSIGNED (NAME).....CERTIFY THAT THE INFORMATION FURNISHED IN PARAGRAPHS 2.

I ACCEPT THAT THE STATE MAY REJECT THE QUOTE OR ACT AGAINST ME SHOULD THIS DECLARATION PROVE TO BE FALSE.

..... Name of bidder Signature Position Date
-------------------------	--------------------	-------------------	---------------

¹"State" means –

- | | |
|---|---|
| a) any national or provincial department, national or provincial public entity or constitutional institution within the meaning of the Public Finance Management Act, 1999 (Act No. 1 of 1999); | c) provincial legislature; |
| b) any municipality or municipal entity; | d) national Assembly or the national Council of provinces; or |
| | e) Parliament. |

²"Shareholder" means a person who owns shares in the company and is actively involved in the management of the enterprise or business and exercises control over the enterprise.