

ZND ST IAL 485/21/22



health

 Department:
 Health
 PROVINCE OF KWAZULU-NATAL

E 3054414.

Specifications Template

Prepared by:

Initial and Surname	Designation	Signature	Date
M Nyuswa	Buyer	<i>M Nyuswa</i>	09-03-2021

Reviewed by Supervisor/Operational Manager:

Initial and Surname	Designation	Signature	Date
<i>I Shabalala</i>	<i>PM</i>	<i>I Shabalala</i>	09-03-21

Item details	Specification
Item description	STOPCOCK, 3-WAY, EXT TUBE, 50.0CM
Size	LENGTH OF TUBING :50 CM INTERNAL DIAMETER OF TUBING: 2.8MM
Colour	
Material	Made from kink resistant tubing with a 3 way stopcock at the proximal end. Manufactured from medical grade plastic. Latex free, non-pyrogenic, DEPH free.
Packaging (unit/box)	Box of 60 <i>Each</i> Individually packed in peel pouch with view paper.
Functionality/performance	
Purpose	Used as an extension between IVI devices. Suitable for use with administration set and an infusion device
Other:	Must be sterile. For single use only

Approved by specifications committee chairperson:

Initial and Surname	Portfolio	Signature	Date
F.Z. Mkhize	<i>Sen</i>	<i>F.Z. Mkhize</i>	09/03/2021