



Department:
Health
PROVINCE OF KWAZULU-NATAL

NEW SPECIFICATION FORM
IALCH SUPPLY CHAIN DEPARTMENT
Postal Address: Private Bag x03 Mayville
Physical Address 800 vusi Mzimela Road Mayville
Tel:0312402094, Fax:0312401050
Email:sifiso.shabalala@ialch.co.za
www.kznhealth.gov.za

SPEC FOR:

F 512 00 56
F 512 00 57
F 512 00 58
F 512 00 59

SPEC NO:	Material no.:
	ZNQ:
ITEM DESCRIPTION	Cannula Left Heart Vent
UNIT OF ISSUE	
SIZES	10Fr 13Fr 16Fr 20Fr
QUALITY STANDARDS	SOUTH AFRICAN BEREU OF STANDADARS (SABS) OR INTERNATIONAL STANDARDS (ISO) APPROVED
ESTIMATED PRICE	

WHAT IS THIS ITEM/PRODUCT USED FOR?

This is left heart vent, used during cardiopulmonary bypass to vent the left ventricle of the heart.

SPECIAL CONDITIONS

- Please submit sample when requested to do so, should you fail to submit, your quotation will be disqualified
- The Department is not compelled to accept lowest price only, evaluation criteria of your bid / quote will be based on Price, Functionality, and as prescribed on Broad Based Black Economic Act and Preferential Procurement Policy

FEATURES EXPECTED FROM THE PRODUCT TO BE EVALUATED (SCOPE)

- Must adapt to 1/4 inch tubing of the bypass circuit.
- Must have a flexible body.

Initial and Surname	Portfolio	Signature	Date
D. Y. J. J. J. J.	Assistant Director		11/11/2020
N. P. Mthembu	Supply		11.11.2020

