



health
 Department:
 Health
 PROVINCE OF KWAZULU-NATAL

E3039102

ZNQ51 JAL308/21/22

Specifications template

Prepared by:

| Initial and Surname | Designation | Signature | Date |
|---------------------|-------------|-----------|------|
| | | | |

Reviewed by Supervisor/Operational Manager:

| Initial and Surname | Designation | Signature | Date |
|---------------------|-------------|-----------|------------|
| E. Jacobs | GM | | 20 08.2021 |

| Item details | Specification |
|---------------------------|--|
| Item description | ARTERIAL FISTULA NEEDLE |
| Size | 16G, 30MM NEEDLE, 300MM TUBE |
| Colour | N/A |
| Material | Lalex-free, BPA-free, DEHP-free |
| Packaging (unit/box) | Unit |
| Functionality/performance | <ul style="list-style-type: none"> Ultra-sharp, siliconized 3-bevel needle for smooth cannulation Ultra-thin wall for optimal blood flow Textured, turnable or fixed, color-coded wings Color-coded occlusion clamp and tapered luer |
| Purpose | Vascular access for hemodialysis patients. |
| Other: | |
| | |

Approved by specifications committee chairperson:

| Initial and Surname | Portfolio | Signature | Date |
|---------------------|-----------|-----------|------------|
| N.P. Mthembu | SCMP | | 23 08.2021 |

